

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL100915\
 Data File : VL026176.D
 Acq On : 9 Oct 2015 13:35
 Operator : SY/MD
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Quant Time: Oct 09 14:35:29 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Oct 09 14:20:41 2015
 QLast Update : Fri Oct 09 14:20:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.67	49	934367	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.34	114	2446465	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.76	117	2821933	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.30	95	2211626	10.62	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.90	62	5987	0.09	ppbv	98
32) 1,1,1-Trichloroethane	7.61	97	14558	0.09	ppbv	99
35) Carbon Tetrachloride	8.18	117	14154	0.09	ppbv	97
38) Trichloroethene	9.07	130	11130	0.10	ppbv	94
52) Tetrachloroethene	12.81	164	11745	0.10	ppbv	93
63) 1,1,2,2-Tetrachloroethane	15.51	83	25591	0.09	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL100915\
Data File : VL026176.D
Acq On : 9 Oct 2015 13:35
Operator : SY/MD
Sample : VSTDIC0.1
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Quant Time: Oct 09 14:35:29 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Oct 09 14:20:41 2015
QLast Update : Fri Oct 09 14:20:41 2015
Response via : Initial Calibration

