

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL100915\
 Data File : VL026177.D
 Acq On : 9 Oct 2015 14:13
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMdadoda
 10/12/2015 1:25:23 PM

Quant Time: Oct 09 15:00:59 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Oct 09 15:00:59 2015
 QLast Update : Fri Oct 09 14:20:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.67	49	914135	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	8.34	114	2403613	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.77	117	2809176	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.31	95	2178325	10.50	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.90	62	1912	0.029	ppbv	92
32) 1,1,1-Trichloroethane	7.62	97	4856	0.030	ppbv	95
35) Carbon Tetrachloride	8.17	117	4758m	0.030	ppbv	
38) Trichloroethene	9.09	130	4500	0.041	ppbv	96
52) Tetrachloroethene	12.82	164	5189	0.045	ppbv #	78
56) 1,1,1,2-Tetrachloroethane	13.80	131	3653m	0.031	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL100915\
Data File : VL026177.D
Acq On : 9 Oct 2015 14:13
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMdadoda
10/12/2015 1:25:23 PM

Quant Time: Oct 09 15:00:59 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Oct
QLast Update : Fri Oct 09 14:20:41 2015
Response via : Initial Calibration

