

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL100915\
 Data File : VL026190.D
 Acq On : 10 Oct 2015 00:14
 Operator : SY/MD
 Sample : G3973-03 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-3

Manual Integrations
 APPROVED

MMdadoda
 10/13/2015 1:17:53 PM

Quant Time: Oct 12 10:44:59 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Oct 0
 QLast Update : Fri Oct 09 16:23:14 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.71	49	274295m	10.00	ppbv	0.05
33) 1,4-Difluorobenzene	8.49	114	606924m	10.00	ppbv	0.16
55) Chlorobenzene-d5	14.01	117	1005482	10.00	ppbv	0.25

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.47	95	3197850m	41.43	ppbv	0.16
Spiked Amount	10.000	Range	65 - 135	Recovery	=	414.30%#

Target Compounds						Ovalue
9) Propene	3.59	41	146026m	9.93	ppbv	
10) Heptane	9.84	43	89917418m	1758.60	ppbv	
13) Ethanol	4.31	45	96453	44.58	ppbv	92
26) Hexane	6.54	57	57331664m	1726.68	ppbv	
30) Cyclohexane	8.56	84	63677518	2012.85	ppbv #	62
36) Benzene	8.24	78	15384260	265.03	ppbv #	77
53) Toluene	11.55	91	4586468	61.32	ppbv	94
58) Ethyl Benzene	14.67	91	7288724	49.91	ppbv	98
59) m/p-Xylene	14.96	91	5382296	45.92	ppbv	98
61) Styrene	15.54	104	372163	7.41	ppbv	93
62) Isopropylbenzene	16.71	105	1043088	6.44	ppbv	96
64) n-propylbenzene	17.66	120	88243	2.02	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL100915\
Data File : VL026190.D
Acq On : 10 Oct 2015 00:14
Operator : SY/MD
Sample : G3973-03 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-3

Manual Integrations
APPROVED

MMdadoda
10/13/2015 1:17:53 PM

Quant Time: Oct 12 10:44:59 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Oct
QLast Update : Fri Oct 09 16:23:14 2015
Response via : Initial Calibration

