

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL100915\
 Data File : VL026200.D
 Acq On : 10 Oct 2015 6:54
 Operator : SY/MD
 Sample : G3925-02DL2 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-2DL2

Manual Integrations
 APPROVED

MMdadoda
 10/12/2015 1:24:58 PM

Quant Time: Oct 12 04:01:26 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Oct 0
 QLast Update : Fri Oct 09 16:23:14 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.64	49	513607	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	8.31	114	1648875	10.00	ppbv	-0.02
55) Chlorobenzene-d5	13.73	117	1905225	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.27	95	1495362	10.22	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.56	51	45655m	0.76	ppbv	
9) Propene	3.60	41	3101	0.11	ppbv #	67
13) Ethanol	4.36	45	1149	0.28	ppbv #	31
15) Acetone	4.62	43	45969	0.69	ppbv #	69
81) 1,2,4-Trichlorobenzene	23.21	180	2589	0.03	ppbv #	12

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL100915\
Data File : VL026200.D
Acq On : 10 Oct 2015 6:54
Operator : SY/MD
Sample : G3925-02DL2 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-2DL2

Manual Integrations
APPROVED

MMdadoda
10/12/2015 1:24:58 PM

Quant Time: Oct 12 04:01:26 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL100915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Oct
QLast Update : Fri Oct 09 16:23:14 2015
Response via : Initial Calibration

