

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101024\
 Data File : VL041636.D
 Acq On : 10 Oct 2024 10:38
 Operator : SY/MD
 Sample : VL1010ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1010ABS01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/11/2024
 Supervised By : Mahesh Dadoda 10/14/2024

Quant Time: Oct 11 05:37:20 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1139395	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.639	114	3120825	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.455	117	3104178m	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.767 95 2267106 9.921 ppbv 0.02
 Spiked Amount 10.000 Range 65 - 135 Recovery = 99.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.706	85	1283214	7.799	ppbv	100
3) Chlorodifluoromethane	2.649	51	1271421	9.200	ppbv	100
4) Chloromethane	2.790	50	701590	9.230	ppbv	100
5) Vinyl Chloride	2.896	62	679675	10.327	ppbv	98
6) Bromomethane	3.095	94	259404	7.922	ppbv	99
7) Chloroethane	3.176	64	244432	9.549	ppbv	99
8) Dichlorotetrafluoroethane	2.835	85	1657160	9.601	ppbv	99
9) Propene	2.671	41	596368	10.230	ppbv	100
10) Heptane	7.565	43	1462473	10.715	ppbv	100
11) Trichlorofluoromethane	3.542	101	1648500	9.415	ppbv	99
12) 1,1,2-Trichlorotrifluo...	4.047	101	1211204	9.551	ppbv	100
13) Ethanol	3.224	45	24686m	5.633	ppbv	
14) Bromoethene	3.343	108	487834	10.145	ppbv	99
15) Acetone	3.446	43	1212615	8.477	ppbv	100
16) 1,3-Butadiene	2.960	39	684097	9.832	ppbv	98
17) tert-Butyl alcohol	3.857	59	1177694	9.362	ppbv	100
18) 1,1-Dichloroethene	3.861	96	541040	10.078	ppbv	96
19) Isopropyl Alcohol	3.555	45	619938	8.466	ppbv #	98
20) Methylene Chloride	3.912	84	438409	8.158	ppbv	93
21) Allyl Chloride	3.976	41	836957	10.095	ppbv	99
22) trans-1,2-Dichloroethene	4.427	96	522871	9.763	ppbv	97
23) Vinyl Acetate	4.610	43	898310	11.417	ppbv	98
24) 1,1-Dichloroethane	4.546	63	1175405	9.953	ppbv	99
25) Ethyl Acetate	5.189	43	2601142	10.129	ppbv	99
26) Hexane	5.195	57	1079483	10.080	ppbv	98
27) Carbon Disulfide	4.105	76	1326465	11.652	ppbv	98
28) Methyl tert-Butyl Ether	4.568	73	671301	10.167	ppbv	100
29) Chloroform	5.259	83	1685416	9.617	ppbv	99
30) Cyclohexane	6.587	84	903279	10.942	ppbv	99
31) cis-1,2-Dichloroethene	5.060	61	1016937	10.002	ppbv	97
32) 1,1,1-Trichloroethane	5.989	97	1657192	10.617	ppbv	98
34) 2-Butanone	4.767	43	1339030	8.998	ppbv	100
35) Carbon Tetrachloride	6.478	117	1724125	11.260	ppbv	99
36) Benzene	6.356	78	2294349	10.586	ppbv	99
37) 1,2-Dichloroethane	5.793	62	1255342	9.983	ppbv	99
38) Trichloroethene	7.279	130	1155435	10.352	ppbv	95
39) 1,2-Dichloropropane	7.063	63	884473	10.031	ppbv	99
40) 1,4-Dioxane	7.272	88	352255	10.598	ppbv #	100
41) Tetrahydrofuran	5.542	42	921578	11.253	ppbv	99
42) Bromodichloromethane	7.237	83	1834395	10.485	ppbv	97
43) Methyl Methacrylate	7.462	69	881765	11.369	ppbv	100
44) 2,2,4-Trimethylpentane	7.314	57	3899661	10.400	ppbv	100
45) t-1,3-Dichloropropene	8.700	75	689956	11.033	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101024\
 Data File : VL041636.D
 Acq On : 10 Oct 2024 10:38
 Operator : SY/MD
 Sample : VL1010ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1010ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/11/2024
 Supervised By :Mahesh Dadoda 10/14/2024

Quant Time: Oct 11 05:37:20 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 17 09:12:21 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.131	75	989234	10.798	ppbv	98
47) 1,1,2-Trichloroethane	8.886	97	944389	10.088	ppbv	99
48) Dibromochloromethane	9.700	129	1544080	10.886	ppbv	100
49) Bromoform	12.397	173	1324724	10.880	ppbv	99
50) 4-Methyl-2-Pentanone	8.169	43	2552765	11.108	ppbv	100
51) 2-Hexanone	9.539	43	2022369	11.634	ppbv	99
52) Tetrachloroethene	10.606	164	845060	10.768	ppbv	97
53) Toluene	9.211	91	2627949	11.422	ppbv	100
54) 1,2-Dibromoethane	9.999	107	1230231	10.671	ppbv	98
56) 1,1,1,2-Tetrachloroethane	11.500	131	1166485	9.625	ppbv	97
57) Chlorobenzene	11.520	112	2044175	9.133	ppbv	99
58) Ethyl Benzene	12.079	91	3540945	10.656	ppbv	99
59) m/p-Xylene	12.362	91	5990335m	20.255	ppbv	
60) o-Xylene	13.050	91	2922422	10.306	ppbv	100
61) Styrene	12.889	104	1385807	11.699	ppbv	100
62) Isopropylbenzene	14.021	105	4217117	9.817	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.037	83	1504875	10.380	ppbv	99
64) n-propylbenzene	14.905	120	1130913	10.193	ppbv	99
65) tert-Butylbenzene	16.082	119	3751383	9.527	ppbv	99
66) Benzyl Chloride	16.317	91	202794	9.906	ppbv	96
67) sec-Butylbenzene	16.625	105	5303112	9.707	ppbv	100
69) p-Isopropyltoluene	16.986	119	4381703	9.908	ppbv	99
70) n-Butylbenzene	17.876	91	4507387	9.730	ppbv	100
71) 2-Chlorotoluene	14.790	91	3141676	9.939	ppbv	99
72) 4-Ethyltoluene	15.185	105	3433852	10.360	ppbv	99
73) 1,3,5-Trimethylbenzene	15.346	105	2950770	10.082	ppbv	95
74) 1,2,4-Trimethylbenzene	16.101	105	3256921	9.507	ppbv	94
75) 1,3-Dichlorobenzene	16.320	146	1962843	9.142	ppbv	99
76) 1,4-Dichlorobenzene	16.465	146	1928062	9.372	ppbv	99
77) 1,2-Dichlorobenzene	17.134	146	1920976	9.063	ppbv	99
78) Hexachloro-1,3-Butadiene	22.625	225	1343127	7.682	ppbv	99
79) Naphthalene	21.374	128	2535281	10.056	ppbv	99
80) Naphthalene,2-methyl-	24.368	142	876186	4.959	ppbv	99
81) 1,2,4-Trichlorobenzene	21.133	180	1347844	8.409	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

