

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101024\
Data File : VL041650.D
Acq On : 10 Oct 2024 20:03
Operator : SY/MD
Sample : CAN 10601
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10601

Quant Time: Oct 11 05:44:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/11/2024
Supervised By :Mahesh Dadoda 10/14/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.173	49	956901	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.635	114	2608928	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.455	117	2284081	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.767	95	1679416	9.988	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101024\
Data File : VL041650.D
Acq On : 10 Oct 2024 20:03
Operator : SY/MD
Sample : CAN 10601
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10601

Quant Time: Oct 11 05:44:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/11/2024
Supervised By : Mahesh Dadoda 10/14/2024

