

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101118\
 Data File : VL032616.D
 Acq On : 11 Oct 2018 22:47
 Operator : SY/AP
 Sample : J5370-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 Q2-SVI-100518DL

Manual Integrations
 APPROVED

MMDadoda
 10/12/2018 2:07:40 PM

Quant Time: Oct 12 03:18:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 08 2018
 QLast Update : Mon Oct 01 06:49:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1458247	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	3251531	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	3399400	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.61	95	2473296	9.65	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.05	41	831130	15.796	ppbv	93
13) Ethanol	3.67	45	56567	3.577	ppbv	91
15) Acetone	3.91	43	4286664	33.489	ppbv	94
17) tert-Butyl alcohol	4.38	59	79731m	4.957	ppbv	
20) Methylene Chloride	4.44	84	21261m	0.343	ppbv	
26) Hexane	5.80	57	31034	0.221	ppbv #	1
27) Carbon Disulfide	4.64	76	50654	0.336	ppbv #	94
34) 2-Butanone	5.35	43	1906133	9.844	ppbv	98
36) Benzene	7.03	78	399134	1.519	ppbv	95
51) 2-Hexanone	10.31	43	272253	1.082	ppbv #	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101118\
Data File : VL032616.D
Acq On : 11 Oct 2018 22:47
Operator : SY/AP
Sample : J5370-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
Q2-SVI-100518DL

Manual Integrations
APPROVED

MMDadoda
10/12/2018 2:07:40 PM

Quant Time: Oct 12 03:18:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct
QLast Update : Mon Oct 01 06:49:26 2018
Response via : Initial Calibration

