

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101123\
 Data File : VL040759.D
 Acq On : 11 Oct 2023 15:10
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/12/2023
 Supervised By : Mahesh Dadoda 10/16/2023

Quant Time: Oct 11 15:42:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101123AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 11 2023
 QLast Update : Wed Oct 11 14:34:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1088922	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.668	114	3007955	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.481	117	2739380	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.789	95	2045737	9.666	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.700%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.951	62	2791m	0.035	ppbv	
32) 1,1,1-Trichloroethane	6.018	97	6449m	0.033	ppbv	
35) Carbon Tetrachloride	6.510	117	5662m	0.027	ppbv	
38) Trichloroethene	7.307	130	4721m	0.056	ppbv	
52) Tetrachloroethene	10.629	164	3376m	0.046	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.063	83	8328m	0.038	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101123\
Data File : VL040759.D
Acq On : 11 Oct 2023 15:10
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Oct 11 15:42:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101123AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Oct 11 14:34:59 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/12/2023
Supervised By : Mahesh Dadoda 10/16/2023

