

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101123\
 Data File : VL040772.D
 Acq On : 12 Oct 2023 1:48
 Operator : SY/MD
 Sample : 04699-13 LOD 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT4-2023

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/13/2023
 Supervised By :Mahesh Dadoda 10/16/2023

Quant Time: Oct 13 01:22:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101123AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2023
 QLast Update : Thu Oct 12 00:21:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.214	49	950745	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.668	114	2703454	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2452867	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1843732	10.215	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.954	62	2341m	0.041	ppbv	
32) 1,1,1-Trichloroethane	6.015	97	5379m	0.035	ppbv	
35) Carbon Tetrachloride	6.504	117	4752m	0.031	ppbv	
38) Trichloroethene	7.311	130	3578m	0.039	ppbv	
52) Tetrachloroethene	10.626	164	3755m	0.044	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101123\
Data File : VL040772.D
Acq On : 12 Oct 2023 1:48
Operator : SY/MD
Sample : 04699-13 LOD 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT4-2023

Quant Time: Oct 13 01:22:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101123AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Oct 12 00:21:08 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/13/2023
Supervised By :Mahesh Dadoda 10/16/2023

