

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101220\
 Data File : VL035777.D
 Acq On : 13 Oct 2020 00:51
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Oct 13 02:24:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Sep 23 07:38:05 2020
 QLast Update : Wed Sep 23 07:38:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	898009	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3326044	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.10	117	3281496	10.00	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.43	95	2374319	9.66	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.60%
Target Compounds						
15) Acetone	3.88	43	15043	0.156	ppbv #	43
20) Methylene Chloride	4.35	84	7279	0.103	ppbv #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101220\
Data File : VL035777.D
Acq On : 13 Oct 2020 00:51
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Oct 13 02:24:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 23 07:38:05 2020
QLast Update : Wed Sep 23 07:38:05 2020
Response via : Initial Calibration

