

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101221\
 Data File : VL037835.D
 Acq On : 12 Oct 2021 18:34
 Operator : SY/AP
 Sample : M4079-05DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P04DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/14/2021
 Supervised By :Mahesh Dadoda 10/15/2021

Quant Time: Oct 14 01:53:36 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1266350	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3806986	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3272984	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2243029	9.80	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	14328m	0.160	ppbv	
15) Acetone	3.85	43	69941m	0.615	ppbv	
20) Methylene Chloride	4.34	84	41032m	0.714	ppbv	
26) Hexane	5.68	57	21566m	0.137	ppbv	
52) Tetrachloroethene	11.22	164	21422m	0.169	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101221\
Data File : VL037835.D
Acq On : 12 Oct 2021 18:34
Operator : SY/AP
Sample : M4079-05DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P04DL

Manual Integrations
APPROVED

Quant Time: Oct 14 01:53:36 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL100421
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 10/14/2021
Supervised By :Mahesh Dadoda 10/15/2021

