

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101221\
 Data File : VL037839.D
 Acq On : 12 Oct 2021 21:12
 Operator : SY/AP
 Sample : M4079-08DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 C0P20DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin
 Yesilyurt

Quant Time: Oct 14 02:11:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1304531	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3831471	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3291051	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2223216	9.66	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	18846m	0.204	ppbv	
13) Ethanol	3.62	45	23629m	3.133	ppbv	
15) Acetone	3.85	43	262827	2.244	ppbv	91
20) Methylene Chloride	4.34	84	46907	0.793	ppbv	97
26) Hexane	5.68	57	27754	0.172	ppbv #	45
34) 2-Butanone	5.26	43	59008m	0.381	ppbv	
52) Tetrachloroethene	11.21	164	16233m	0.127	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101221\
Data File : VL037839.D
Acq On : 12 Oct 2021 21:12
Operator : SY/AP
Sample : M4079-08DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P20DL

Quant Time: Oct 14 02:11:37 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL100421
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

10/14/2021
Supervised By :Mahesh
Dadoda

10/15/2021

