

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101221\
 Data File : VL037850.D
 Acq On : 13 Oct 2021 10:39
 Operator : SY/AP
 Sample : OC101221 CAN 1065
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 QC101221 CAN 1065

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/14/2021
 Supervised By :Mahesh Dadoda 10/15/2021

Quant Time: Oct 14 03:12:04 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1225836	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3532083	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	2909383	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	1993603	9.80	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	14627m	0.169	ppbv	
13) Ethanol	3.62	45	44405m	6.265	ppbv	
15) Acetone	3.85	43	96184m	0.874	ppbv	
20) Methylene Chloride	4.34	84	304087	5.469	ppbv	97
26) Hexane	5.68	57	152624	1.004	ppbv #	44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101221\
Data File : VL037850.D
Acq On : 13 Oct 2021 10:39
Operator : SY/AP
Sample : OC101221 CAN 1065
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC101221 CAN 1065

Quant Time: Oct 14 03:12:04 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/14/2021
Supervised By :Mahesh Dadoda 10/15/2021

