

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101419\
 Data File : VL034293.D
 Acq On : 14 Oct 2019 22:37
 Operator : SY/AP
 Sample : K5302-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

MMDadoda
 10/16/2019 12:59:39 AM

Quant Time: Oct 15 07:42:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep 18 2019
 QLast Update : Wed Sep 18 12:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1006796	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.21	114	1666659	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.15	117	1642962	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.49	95	1462729	10.25	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	7677m	0.544	ppbv	
15) Acetone	3.85	43	50909	0.377	ppbv	90
20) Methylene Chloride	4.35	84	8986m	0.231	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101419\
Data File : VL034293.D
Acq On : 14 Oct 2019 22:37
Operator : SY/AP
Sample : K5302-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QA1DL

Manual Integrations
APPROVED

MMDadoda
10/16/2019 12:59:39 AM

Quant Time: Oct 15 07:42:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep
QLast Update : Wed Sep 18 12:53:31 2019
Response via : Initial Calibration

