

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101419\
 Data File : VL034295.D
 Acq On : 14 Oct 2019 23:58
 Operator : SY/AP
 Sample : OC101119 CAN 10296
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC101119 CAN 10296

Manual Integrations
 APPROVED

MMDadoda
 10/16/2019 12:59:41 AM

Quant Time: Oct 15 07:44:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Sep 18
 QLast Update : Wed Sep 18 12:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	948908	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.21	114	1580233	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.15	117	1574954	10.00	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.49	95	1424583	10.41	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.10%
Target Compounds						
15) Acetone	3.86	43	33132	0.261	ppbv #	86
20) Methylene Chloride	4.35	84	18626m	0.508	ppbv	
26) Hexane	5.71	57	12290	0.123	ppbv #	41

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101419\
Data File : VL034295.D
Acq On : 14 Oct 2019 23:58
Operator : SY/AP
Sample : OC101119 CAN 10296
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC101119 CAN 10296

Manual Integrations
APPROVED

MMDadoda
10/16/2019 12:59:41 AM

Quant Time: Oct 15 07:44:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL091719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Sep
QLast Update : Wed Sep 18 12:53:31 2019
Response via : Initial Calibration

