


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL101419\
Data File : VL034277.D
Acq On : 14 Oct 2019 8:42
Operator : SY/AP
Sample : BFB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091719AIR.M
Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Sep 18 12:53:31 2019
Last Update : Wed Sep 18 12:53:31 2019



AutoFind: Scans 3638, 3639, 3640; Background Corrected with Scan 3615

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	26.7	181162	PASS
75	95	30	66	60.2	407616	PASS
95	95	100	100	100.0	677269	PASS
96	95	5	9	6.4	43586	PASS
173	174	0.00	2	0.9	3580	PASS
174	95	50	120	60.5	409685	PASS
175	174	4	9	7.2	29301	PASS
176	174	93	101	97.7	400448	PASS
177	176	5	9	6.1	24522	PASS