

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041664.D
 Acq On : 14 Oct 2024 12:20
 Operator : SY/MD
 Sample : CAN 10258
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10258

Quant Time: Oct 15 02:40:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.201	49	1000184	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	6.667	114	2684060	10.000	ppbv	0.04
55) Chlorobenzene-d5	11.487	117	2368428	10.000	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.799	95	1743192	9.998	ppbv	0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
Data File : VL041664.D
Acq On : 14 Oct 2024 12:20
Operator : SY/MD
Sample : CAN 10258
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10258

Quant Time: Oct 15 02:40:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration

