

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
 Data File : VL041670.D
 Acq On : 14 Oct 2024 16:20
 Operator : SY/MD
 Sample : P4368-15 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL-AIR-03-QT4-2024

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/15/2024
 Supervised By :Mahesh Dadoda 10/15/2024

Quant Time: Oct 15 04:25:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 17 09:12:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	958409	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	6.664	114	2608385	10.000	ppbv	0.04
55) Chlorobenzene-d5	11.484	117	2337623	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.793	95	1714652	9.964	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.919	62	1836m	0.033	ppbv	
32) 1,1,1-Trichloroethane	6.015	97	4062m	0.031	ppbv	
35) Carbon Tetrachloride	6.507	117	3862m	0.030	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.076	83	3516m	0.032	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101424\
Data File : VL041670.D
Acq On : 14 Oct 2024 16:20
Operator : SY/MD
Sample : P4368-15 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL-AIR-03-QT4-2024

Quant Time: Oct 15 04:25:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 17 09:12:21 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/15/2024
Supervised By : Mahesh Dadoda 10/15/2024

