

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL101718\
 Data File : VL032648.D
 Acq On : 17 Oct 2018 19:32
 Operator : SY/AP
 Sample : J5500-02DL2 1000X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS04-MIVIDL2

Manual Integrations
 APPROVED

MMDadoda
 10/19/2018 4:36:42 PM

Quant Time: Oct 18 10:49:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Oct 18 10:49:09 2018
 QLast Update : Wed Oct 17 08:55:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.78	49	1179675	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.31	114	2537469	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.26	117	2426287	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.60	95	1875137	10.26	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.06	41	19516m	0.54	ppbv	
15) Acetone	3.94	43	63800	0.61	ppbv	94
18) 1,1-Dichloroethene	4.37	96	12257	0.24	ppbv #	69
29) Chloroform	5.86	83	3856110	19.66	ppbv	100
35) Carbon Tetrachloride	7.15	117	1024101	5.21	ppbv	98
38) Trichloroethene	7.98	130	9110m	0.12	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL101718\
Data File : VL032648.D
Acq On : 17 Oct 2018 19:32
Operator : SY/AP
Sample : J5500-02DL2 1000X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS04-MIVIDL2

Manual Integrations
APPROVED

MMDadoda
10/19/2018 4:36:42 PM

Quant Time: Oct 18 10:49:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL101718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Oct
QLast Update : Wed Oct 17 08:55:11 2018
Response via : Initial Calibration

