

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL101718\
 Data File : VL032634.D
 Acq On : 17 Oct 2018 7:15
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL101718

Quant Time: Oct 17 16:28:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Oct 17 08:55:11 2018
 QLast Update : Wed Oct 17 08:55:11 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.986	0.818	17.0	110	0.00
3	Chlorodifluoromethane	0.837	0.838	-0.1	103	0.00
4	Chloromethane	0.457	0.474	-3.7	103	0.00
5 T	Vinyl Chloride	0.460	0.481	-4.6	102	0.00
6 T	Bromomethane	0.291	0.287	1.4	99	0.00
7	Chloroethane	0.180	0.182	-1.1	101	0.00
8 T	Dichlorotetrafluoroethane	1.127	1.098	2.6	101	0.00
9 T	Propene	0.307	0.355	-15.6	106	0.00
10 T	Heptane	1.149	1.306	-13.7	98	0.00
11 T	Trichlorofluoromethane	1.285	1.246	3.0	106	0.00
12 T	1,1,2-Trichlorotrifluoroeth	0.941	0.869	7.7	100	0.00
13	Ethanol	0.095	0.093	2.1	104	0.00
14 T	Bromoethene	0.344	0.350	-1.7	100	0.00
15 T	Acetone	0.890	0.814	8.5	98	0.00
16 T	1,3-Butadiene	0.330	0.340	-3.0	102	0.00
17	tert-Butyl alcohol	0.081	0.085	-4.9	104	0.00
18 T	1,1-Dichloroethene	0.424	0.417	1.7	102	0.00
19 T	Isopropyl Alcohol	0.628	0.688	-9.6	114	0.00
20 T	Methylene Chloride	0.399	0.359	10.0	99	0.00
21 T	Allyl Chloride	0.668	0.704	-5.4	102	0.00
22 T	trans-1,2-Dichloroethene	0.490	0.482	1.6	100	0.00
23 T	Vinyl Acetate	1.416	1.637	-15.6	101	0.00
24 T	1,1-Dichloroethane	1.313	1.314	-0.1	100	0.00
25 T	Ethyl Acetate	2.106	2.100	0.3	97	0.00
26 T	Hexane	0.948	0.983	-3.7	97	0.00
27 T	Carbon Disulfide	1.146	1.118	2.4	99	0.00
28 T	Methyl tert-Butyl Ether	1.192	1.361	-14.2	102	0.00
29 T	Chloroform	1.662	1.635	1.6	99	0.00
30 T	Cyclohexane	0.748	0.860	-15.0	101	0.00
31 T	cis-1,2-Dichloroethene	0.911	1.029	-13.0	99	0.00
32 T	1,1,1-Trichloroethane	1.658	1.602	3.4	100	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
34 T	2-Butanone	0.555	0.611	-10.1	100	0.00
35 T	Carbon Tetrachloride	0.775	0.712	8.1	99	0.00
36 T	Benzene	0.811	0.883	-8.9	100	0.00
37 T	1,2-Dichloroethane	0.556	0.547	1.6	99	0.00
38 T	Trichloroethene	0.305	0.325	-6.6	100	0.00
39 T	1,2-Dichloropropane	0.330	0.344	-4.2	101	0.00
40 T	1,4-Dioxane	0.114	0.115	-0.9	89	0.00
41 T	Tetrahydrofuran	0.294	0.346	-17.7	100	0.00
42 T	Bromodichloromethane	0.758	0.756	0.3	99	0.00
43	Methyl Methacrylate	0.314	0.370	-17.8	101	0.00
44 T	2,2,4-Trimethylpentane	1.437	1.540	-7.2	100	0.00
45 T	t-1,3-Dichloropropene	0.411	0.516	-25.5	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL101718\
 Data File : VL032634.D
 Acq On : 17 Oct 2018 7:15
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL101718

Quant Time: Oct 17 16:28:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Oct 17 08:55:11 2018
 QLast Update : Wed Oct 17 08:55:11 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.468	0.563	-20.3	99	0.00
47 T	1,1,2-Trichloroethane	0.394	0.392	0.5	99	0.00
48 T	Dibromochloromethane	0.669	0.682	-1.9	98	0.00
49 T	Bromoform	0.506	0.516	-2.0	98	0.00
50 T	4-Methyl-2-Pentanone	0.817	0.892	-9.2	97	0.00
51 T	2-Hexanone	0.642	0.784	-22.1	100	0.00
52 T	Tetrachloroethene	0.267	0.283	-6.0	100	0.00
53 T	Toluene	0.924	1.101	-19.2	100	0.00
54 T	1,2-Dibromoethane	0.564	0.610	-8.2	100	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
56	1,1,1,2-Tetrachloroethane	0.459	0.436	5.0	98	0.00
57 T	Chlorobenzene	0.761	0.749	1.6	99	0.00
58 T	Ethyl Benzene	1.123	1.342	-19.5	99	0.00
59 T	m/p-Xylene	1.003	1.103	-10.0	98	0.00
60 T	o-Xylene	0.999	1.091	-9.2	99	0.00
61 T	Styrene	0.605	0.783	-29.4	99	0.00
62	Isopropylbenzene	1.423	1.529	-7.4	98	0.00
63 T	1,1,2,2-Tetrachloroethane	0.882	0.810	8.2	98	0.00
64	n-propylbenzene	0.361	0.393	-8.9	99	0.00
65	tert-Butylbenzene	1.208	1.297	-7.4	98	0.00
66 T	Benzyl Chloride	0.681	0.777	-14.1	98	0.00
67	sec-Butylbenzene	1.817	1.925	-5.9	99	0.00
68 S	1-Bromo-4-Fluorobenzene	0.753	0.764	-1.5	97	0.00
69	p-Isopropyltoluene	1.376	1.529	-11.1	98	0.00
70	n-Butylbenzene	1.516	1.656	-9.2	99	0.00
71	2-Chlorotoluene	1.052	1.204	-14.4	99	0.00
72 T	4-Ethyltoluene	1.080	1.257	-16.4	98	0.00
73 T	1,3,5-Trimethylbenzene	1.074	1.194	-11.2	98	0.00
74 T	1,2,4-Trimethylbenzene	1.183	1.208	-2.1	97	0.00
75 T	1,3-Dichlorobenzene	0.687	0.662	3.6	98	0.00
76 T	1,4-Dichlorobenzene	0.626	0.650	-3.8	97	0.00
77 T	1,2-Dichlorobenzene	0.661	0.661	0.0	99	0.00
78 T	Hexachloro-1,3-Butadiene	0.370	0.393	-6.2	108	0.00
79 T	Naphthalene	0.804	1.005	-25.0	98	0.00
80 T	Naphthalene,2-methyl-	0.156	0.297	-90.4#	106	0.00
81 T	1,2,4-Trichlorobenzene	0.399	0.476	-19.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0