

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101719\
 Data File : VL034326.D
 Acq On : 17 Oct 2019 21:21
 Operator : SY/AP
 Sample : K5409-01DL2 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL2

Manual Integrations
 APPROVED

MMDadoda
 10/22/2019 2:55:22 PM

Quant Time: Oct 18 05:32:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL101619AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 18 05:32:54 2019
 QLast Update : Thu Oct 17 05:12:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	935465	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.21	114	1957475	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.16	117	2002392	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.49	95	1660893	9.79	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	39640	0.816	ppbv	94
15) Acetone	3.84	43	442477	3.307	ppbv	99
20) Methylene Chloride	4.35	84	11623	0.240	ppbv	95
27) Carbon Disulfide	4.56	76	9524m	0.095	ppbv	
31) cis-1,2-Dichloroethene	5.57	61	19690	0.202	ppbv #	83
34) 2-Butanone	5.28	43	32717m	0.204	ppbv	
38) Trichloroethene	7.87	130	13669m	0.181	ppbv	
52) Tetrachloroethene	11.29	164	1245958	16.925	ppbv	96
53) Toluene	9.86	91	26360	0.127	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101719\
Data File : VL034326.D
Acq On : 17 Oct 2019 21:21
Operator : SY/AP
Sample : K5409-01DL2 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL2

Manual Integrations
APPROVED

MMDadoda
10/22/2019 2:55:22 PM

Quant Time: Oct 18 05:32:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL101619AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 17 05:12:37 2019
Response via : Initial Calibration

