

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102015\
 Data File : VL026210.D
 Acq On : 20 Oct 2015 13:10
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Quant Time: Oct 20 14:20:21 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Oct 20 12:43:06 2015
 QLast Update : Tue Oct 20 12:43:06 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.63	49	557419	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.30	114	1641175	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.71	117	1962061	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.26	95	1558294	10.34	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.87	62	4037	0.11	ppbv	97
32) 1,1,1-Trichloroethane	7.57	97	12529	0.13	ppbv	98
35) Carbon Tetrachloride	8.13	117	13985	0.13	ppbv	97
38) Trichloroethene	9.04	130	9247	0.13	ppbv	93
52) Tetrachloroethene	12.77	164	11145	0.14	ppbv	95
63) 1,1,2,2-Tetrachloroethane	15.46	83	18843	0.11	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102015\
Data File : VL026210.D
Acq On : 20 Oct 2015 13:10
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Quant Time: Oct 20 14:20:21 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Oct 20 12:43:06 2015
QLast Update : Tue Oct 20 12:43:06 2015
Response via : Initial Calibration

