

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102015\
 Data File : VL026213.D
 Acq On : 20 Oct 2015 15:21
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Oct 21 02:24:53 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Oct 20 20:24:48 2015
 QLast Update : Tue Oct 20 20:24:48 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.63	49	568907	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.31	114	1656488	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.72	117	1945475	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.26	95	1585608	10.57	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	4.63	43	28397	0.34	ppbv	# 37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102015\
Data File : VL026213.D
Acq On : 20 Oct 2015 15:21
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Oct 21 02:24:53 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Oct 20 20:24:48 2015
QLast Update : Tue Oct 20 20:24:48 2015
Response via : Initial Calibration

