

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102115\
 Data File : VL026236.D
 Acq On : 21 Oct 2015 13:56
 Operator : SY/MD
 Sample : G4112-01 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SUBSLAB-233

Quant Time: Oct 22 09:25:37 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Oct 21 01:20:55 2015
 QLast Update : Wed Oct 21 01:20:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.62	49	503616	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	8.29	114	1474935	10.00	ppbv	-0.02
55) Chlorobenzene-d5	13.71	117	1775301	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.25	95	1408951	10.29	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.62	85	12562	0.12	ppbv	93
3) Chlorodifluoromethane	3.55	51	20029	0.31	ppbv	97
13) Ethanol	4.33	45	30737	6.87	ppbv	88
15) Acetone	4.60	43	66103	0.89	ppbv #	80
17) tert-Butyl alcohol	5.17	59	9578	0.19	ppbv #	87
26) Hexane	6.63	57	13107	0.19	ppbv #	55
34) 2-Butanone	6.21	43	9498	0.11	ppbv	94
38) Trichloroethene	9.03	130	4958	0.07	ppbv	95
41) Tetrahydrofuran	7.15	42	8360	0.17	ppbv	95
52) Tetrachloroethene	12.76	164	1519669	17.64	ppbv	99
53) Toluene	11.20	91	67285	0.33	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102115\
Data File : VL026236.D
Acq On : 21 Oct 2015 13:56
Operator : SY/MD
Sample : G4112-01 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SUBSLAB-233

Quant Time: Oct 22 09:25:37 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Oct 21 01:20:55 2015
QLast Update : Wed Oct 21 01:20:55 2015
Response via : Initial Calibration

