

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102115\
 Data File : VL026237.D
 Acq On : 21 Oct 2015 14:35
 Operator : SY/MD
 Sample : G4112-02DL 100X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 RB-SUBSLAB-239DL

Manual Integrations
 APPROVED

MMDadoda
 10/22/2015 4:35:50 PM

Quant Time: Oct 22 10:18:44 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Oct 22 10:18:44 2015
 QLast Update : Wed Oct 21 01:20:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.62	49	480465	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	8.29	114	1339649	10.00	ppbv	-0.02
55) Chlorobenzene-d5	13.71	117	1683037	10.00	ppbv	-0.02

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	16.25	95	1339851	10.32	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.62	85	28280	0.28	ppbv	98
3) Chlorodifluoromethane	3.56	51	56086m	0.92	ppbv	
13) Ethanol	4.34	45	19755	4.62	ppbv	95
15) Acetone	4.61	43	50588	0.72	ppbv #	75
26) Hexane	6.63	57	7655	0.12	ppbv #	53
34) 2-Butanone	6.21	43	7682	0.10	ppbv	99
38) Trichloroethene	9.03	130	9961	0.15	ppbv	96
52) Tetrachloroethene	12.76	164	1196519	15.29	ppbv	99
53) Toluene	11.21	91	35699	0.19	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102115\
Data File : VL026237.D
Acq On : 21 Oct 2015 14:35
Operator : SY/MD
Sample : G4112-02DL 100X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
RB-SUBSLAB-239DL

Manual Integrations
APPROVED

MMDadoda
10/22/2015 4:35:50 PM

Quant Time: Oct 22 10:18:44 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102015AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Oct
QLast Update : Wed Oct 21 01:20:55 2015
Response via : Initial Calibration

