

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_L\DATA\VL102419\
 Data File : VL034347.D
 Acq On : 24 Oct 2019 15:37
 Operator : SY/AP
 Sample : K5477-01DL2 40X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV5DL2

Manual Integrations
 APPROVED

MMDadoda
 10/28/2019 9:43:17 AM

Quant Time: Oct 27 19:42:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101619AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 1
 QLast Update : Thu Oct 17 05:12:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	951382	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.21	114	2059621	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.17	117	1946552	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.50	95	1544011	9.37	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
22) trans-1,2-Dichloroethene	4.89	96	39498	0.887	ppbv #	90
29) Chloroform	5.76	83	179925	1.162	ppbv	96
31) cis-1,2-Dichloroethene	5.56	61	3882480	39.068	ppbv #	80
36) Benzene	6.92	78	48815	0.247	ppbv #	91
38) Trichloroethene	7.87	130	3424379	43.183	ppbv	96
44) 2,2,4-Trimethylpentane	7.91	57	103580m	0.283	ppbv	
52) Tetrachloroethene	11.42	164	37525040	484.444	ppbv	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_L\DATA\VL102419\
Data File : VL034347.D
Acq On : 24 Oct 2019 15:37
Operator : SY/AP
Sample : K5477-01DL2 40X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV5DL2

Manual Integrations
APPROVED

MMDadoda
10/28/2019 9:43:17 AM

Quant Time: Oct 27 19:42:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101619AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 17 05:12:37 2019
Response via : Initial Calibration

