

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102422\
 Data File : VL039766.D
 Acq On : 24 Oct 2022 19:03
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/26/2022
 Supervised By :Mahesh Dadoda 10/31/2022

Quant Time: Oct 24 19:47:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Oct 24 17:14:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	903651	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.655	114	2435403	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2067180	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	1497044	10.202	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.000%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.948	62	1921m	0.034	ppbv	
32) 1,1,1-Trichloroethane	6.018	97	3561m	0.029	ppbv	
35) Carbon Tetrachloride	6.494	117	4597m	0.036	ppbv	
38) Trichloroethene	7.275	130	3215m	0.033	ppbv	
52) Tetrachloroethene	10.597	164	2423m	0.037	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.056	83	3238m	0.042	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102422\
Data File : VL039766.D
Acq On : 24 Oct 2022 19:03
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/26/2022
Supervised By :Mahesh Dadoda 10/31/2022

Quant Time: Oct 24 19:47:54 2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Mon Oct 24 17:14:55 2022

Response via : Initial Calibration

