

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102422\
 Data File : VL039768.D
 Acq On : 24 Oct 2022 22:14
 Operator : SY/AP
 Sample : VL1024ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1024ABL01

Quant Time: Oct 24 23:25:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Oct 24 20:00:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	913952	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.652	114	2396403	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	2052324	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.767	95	1433174	9.817	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102422\
Data File : VL039768.D
Acq On : 24 Oct 2022 22:14
Operator : SY/AP
Sample : VL1024ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1024ABL01

Quant Time: Oct 24 23:25:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Oct 24 20:00:55 2022
Response via : Initial Calibration

