

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_L\DATA\VL102519\
 Data File : VL034367.D
 Acq On : 25 Oct 2019 16:59
 Operator : SY/AP
 Sample : VBLK
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VBLK

Manual Integrations
 APPROVED

MMDadoda
 10/31/2019 2:12:59 AM

Quant Time: Oct 26 00:16:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101619AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 1
 QLast Update : Thu Oct 17 05:12:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	871463	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	1715068	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.14	117	1711260	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.48	95	1474325	10.17	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	3.02	85	15425	0.103	ppbv #	80
9) Propene	3.00	41	7822m	0.173	ppbv	
13) Ethanol	3.60	45	47856	4.720	ppbv	99
15) Acetone	3.84	43	110913	0.890	ppbv	94
19) Isopropyl Alcohol	3.99	45	8102	0.117	ppbv #	1
20) Methylene Chloride	4.34	84	229254	5.087	ppbv	95
26) Hexane	5.70	57	74858	0.835	ppbv #	41
31) cis-1,2-Dichloroethene	5.56	61	8467	0.093	ppbv	97
34) 2-Butanone	5.26	43	13894m	0.099	ppbv	
38) Trichloroethene	7.87	130	8841m	0.134	ppbv	
52) Tetrachloroethene	11.28	164	319884	4.959	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_L\DATA\VL102519\
Data File : VL034367.D
Acq On : 25 Oct 2019 16:59
Operator : SY/AP
Sample : VBLK
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VBLK

Manual Integrations
APPROVED

MMDadoda
10/31/2019 2:12:59 AM

Quant Time: Oct 26 00:16:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101619AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 17 05:12:37 2019
Response via : Initial Calibration

