

Data Path : \\74.0.250.170\Terastorage\voasrv\HPCHEM1\MSVOA_L\Data\VL102519\
 Data File : VL034359.D
 Acq On : 25 Oct 2019 10:14
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 25 23:49:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101619AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 17 05:12:37 2019
 QLast Update : Thu Oct 17 05:12:37 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	1.724	1.783	-3.4	111	0.00
3	Chlorodifluoromethane	1.172	1.059	9.6	71	0.00
4	Chloromethane	0.693	0.650	6.2	73	0.00
5 T	Vinyl Chloride	0.661	0.605	8.5	75	0.00
6 T	Bromomethane	0.324	0.318	1.9	75	0.00
7	Chloroethane	0.233	0.227	2.6	73	0.00
8 T	Dichlorotetrafluoroethane	1.494	1.408	5.8	78	0.00
9 T	Propene	0.520	0.520	0.0	74	0.00
10 T	Heptane	1.334	1.331	0.2	73	0.00
11 T	Trichlorofluoromethane	1.637	1.539	6.0	76	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.072	1.032	3.7	76	0.00
13	Ethanol	0.116	0.087	25.0	60	0.00
14 T	Bromoethene	0.422	0.445	-5.5	77	0.00
15 T	Acetone	1.430	1.141	20.2	74	0.00
16 T	1,3-Butadiene	0.581	0.573	1.4	74	0.00
17	tert-Butyl alcohol	0.913	0.830	9.1	59	0.00
18 T	1,1-Dichloroethene	0.470	0.477	-1.5	78	0.00
19 T	Isopropyl Alcohol	0.798	0.652	18.3	54	0.00
20 T	Methylene Chloride	0.517	0.422	18.4	77	0.00
21 T	Allyl Chloride	0.791	0.801	-1.3	74	0.00
22 T	trans-1,2-Dichloroethene	0.468	0.487	-4.1	77	0.00
23 T	Vinyl Acetate	1.786	1.732	3.0	70	0.00
24 T	1,1-Dichloroethane	1.326	1.255	5.4	73	0.00
25 T	Ethyl Acetate	2.415	2.262	6.3	74	0.00
26 T	Hexane	1.029	0.994	3.4	76	0.00
27 T	Carbon Disulfide	1.069	1.191	-11.4	75	0.00
28 T	Methyl tert-Butyl Ether	1.558	1.702	-9.2	81	0.00
29 T	Chloroform	1.628	1.575	3.3	76	0.00
30 T	Cyclohexane	0.757	0.811	-7.1	78	0.00
31 T	cis-1,2-Dichloroethene	1.045	1.071	-2.5	75	0.00
32 T	1,1,1-Trichloroethane	1.583	1.540	2.7	75	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
34 T	2-Butanone	0.817	0.713	12.7	73	0.00
35 T	Carbon Tetrachloride	0.871	0.713	18.1	71	0.00
36 T	Benzene	0.959	0.908	5.3	79	0.00
37 T	1,2-Dichloroethane	0.715	0.603	15.7	71	0.00
38 T	Trichloroethene	0.385	0.345	10.4	77	0.00
39 T	1,2-Dichloropropane	0.413	0.377	8.7	76	0.00
40 T	1,4-Dioxane	0.145	0.122	15.9	65	0.00
41 T	Tetrahydrofuran	0.454	0.420	7.5	73	0.00
42 T	Bromodichloromethane	0.876	0.805	8.1	76	-0.01
43	Methyl Methacrylate	0.397	0.381	4.0	76	0.00
44 T	2,2,4-Trimethylpentane	1.776	1.565	11.9	75	0.00
45 T	t-1,3-Dichloropropene	0.483	0.507	-5.0	74	0.00

Data Path : \\74.0.250.170\Terastorage\voasrv\HPCHEM1\MSVOA_L\Data\VL102519\
 Data File : VL034359.D
 Acq On : 25 Oct 2019 10:14
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 25 23:49:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL101619AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 17 05:12:37 2019
 QLast Update : Thu Oct 17 05:12:37 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.555	0.575	-3.6	75	0.00
47 T	1,1,2-Trichloroethane	0.418	0.374	10.5	76	0.00
48 T	Dibromochloromethane	0.686	0.655	4.5	74	0.00
49 T	Bromoform	0.608	0.575	5.4	71	0.00
50 T	4-Methyl-2-Pentanone	1.170	1.060	9.4	73	-0.01
51 T	2-Hexanone	0.959	0.892	7.0	73	0.00
52 T	Tetrachloroethene	0.376	0.342	9.0	78	0.00
53 T	Toluene	1.059	1.050	0.8	79	0.00
54 T	1,2-Dibromoethane	0.620	0.570	8.1	77	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
56	1,1,1,2-Tetrachloroethane	0.461	0.417	9.5	74	0.00
57 T	Chlorobenzene	0.834	0.765	8.3	77	-0.01
58 T	Ethyl Benzene	1.398	1.415	-1.2	79	-0.01
59 T	m/p-Xylene	1.211	1.156	4.5	77	-0.02
60 T	o-Xylene	1.206	1.147	4.9	77	0.00
61 T	Styrene	0.849	0.871	-2.6	77	-0.01
62	Isopropylbenzene	1.592	1.508	5.3	77	-0.01
63 T	1,1,2,2-Tetrachloroethane	0.969	0.776	19.9	73	0.00
64	n-propylbenzene	0.409	0.391	4.4	76	0.00
65	tert-Butylbenzene	1.411	1.313	6.9	75	-0.01
66 T	Benzyl Chloride	0.557	0.520	6.6	59	-0.01
67	sec-Butylbenzene	1.986	1.823	8.2	76	0.00
68 S	1-Bromo-4-Fluorobenzene	0.847	0.835	1.4	79	0.00
69	p-Isopropyltoluene	1.619	1.509	6.8	75	0.00
70	n-Butylbenzene	1.761	1.617	8.2	74	0.00
71	2-Chlorotoluene	1.234	1.186	3.9	75	-0.01
72 T	4-Ethyltoluene	1.356	1.325	2.3	76	-0.01
73 T	1,3,5-Trimethylbenzene	1.343	1.258	6.3	75	0.00
74 T	1,2,4-Trimethylbenzene	1.401	1.276	8.9	74	0.00
75 T	1,3-Dichlorobenzene	0.841	0.728	13.4	73	-0.01
76 T	1,4-Dichlorobenzene	0.822	0.728	11.4	72	0.00
77 T	1,2-Dichlorobenzene	0.825	0.710	13.9	72	0.00
78 T	Hexachloro-1,3-Butadiene	0.682	0.693	-1.6	93	-0.01
79 T	Naphthalene	1.023	1.270	-24.1	88	-0.02
80 T	Naphthalene,2-methyl-	0.296	0.633	-113.9#	132	0.00
81 T	1,2,4-Trichlorobenzene	0.668	0.741	-10.9	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0