

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
 Data File : VL039793.D
 Acq On : 25 Oct 2022 18:32
 Operator : SY/AP
 Sample : N5227-12DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/28/2022
 Supervised By :Mahesh Dadoda 10/31/2022

Quant Time: Oct 27 17:29:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Oct 24 20:00:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.202	49	871212	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.658	114	2231801	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	1894429	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.774	95	1356016	10.062	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.600%
Target Compounds						
						Qvalue
15) Acetone	3.504	43	44403	0.427	ppbv	92
20) Methylene Chloride	3.951	84	81534	2.059	ppbv	91
26) Hexane	5.230	57	43276m	0.512	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
Data File : VL039793.D
Acq On : 25 Oct 2022 18:32
Operator : SY/AP
Sample : N5227-12DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LDL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/28/2022
Supervised By : Mahesh Dadoda 10/31/2022

Quant Time: Oct 27 17:29:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Oct 24 20:00:55 2022
Response via : Initial Calibration

