

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
 Data File : VL039795.D
 Acq On : 25 Oct 2022 19:52
 Operator : SY/AP
 Sample : N5227-13DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL

Quant Time: Oct 28 08:39:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Oct 24 20:00:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	868892	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.661	114	2238991	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	1910329	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.773	95	653866	4.812	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	48.100%#
Target Compounds						
					Qvalue	
9) Propene	2.716	41	55721	1.243	ppbv	98
15) Acetone	3.504	43	97158	0.936	ppbv	97
20) Methylene Chloride	3.951	84	53820	1.363	ppbv #	88
26) Hexane	5.224	57	30794	0.365	ppbv #	34

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
Data File : VL039795.D
Acq On : 25 Oct 2022 19:52
Operator : SY/AP
Sample : N5227-13DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL

Quant Time: Oct 28 08:39:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Oct 24 20:00:55 2022
Response via : Initial Calibration

