

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
 Data File : VL039799.D
 Acq On : 25 Oct 2022 22:37
 Operator : SY/AP
 Sample : N5227-08DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 HDL

Quant Time: Oct 27 17:31:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Oct 24 20:00:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	838306	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.661	114	2361917	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	1986112	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1397434	9.891	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.900%
Target Compounds						
					Qvalue	
9) Propene	2.709	41	476391	11.010	ppbv	91
15) Acetone	3.497	43	173898	1.737	ppbv #	83
20) Methylene Chloride	3.947	84	11297	0.297	ppbv #	87
26) Hexane	5.227	57	46824	0.575	ppbv #	34
36) Benzene	6.378	78	145270	0.799	ppbv #	94
53) Toluene	9.224	91	83007	0.436	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
Data File : VL039799.D
Acq On : 25 Oct 2022 22:37
Operator : SY/AP
Sample : N5227-08DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
HDL

Quant Time: Oct 27 17:31:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Oct 24 20:00:55 2022
Response via : Initial Calibration

