

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
 Data File : VL039803.D
 Acq On : 26 Oct 2022 1:20
 Operator : SY/AP
 Sample : N5227-10DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 JDL

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/31/2022
 Supervised By :Mahesh Dadoda 10/31/2022

Quant Time: Oct 28 08:40:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Oct 24 20:00:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	843503	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.664	114	2159904	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	1881293	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1278440	9.553	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.500%
Target Compounds						
						Qvalue
15) Acetone	3.504	43	64192	0.637	ppbv	97
20) Methylene Chloride	3.960	84	19160m	0.500	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
Data File : VL039803.D
Acq On : 26 Oct 2022 1:20
Operator : SY/AP
Sample : N5227-10DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
JDL

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/31/2022
Supervised By :Mahesh Dadoda 10/31/2022

Quant Time: Oct 28 08:40:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Mon Oct 24 20:00:55 2022
Response via : Initial Calibration

