

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
 Data File : VL039806.D
 Acq On : 26 Oct 2022 9:01
 Operator : SY/AP
 Sample : QC102422 CAN 10279
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC102422 CAN 10279

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/31/2022
 Supervised By :Mahesh Dadoda 10/31/2022

Quant Time: Oct 28 08:41:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 2022
 QLast Update : Mon Oct 24 20:00:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	821084	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.668	114	2152730	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.475	117	1817679	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.786	95	1254438	9.701	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.000%
Target Compounds						
					Qvalue	
15) Acetone	3.520	43	11939	0.122	ppbv #	78
20) Methylene Chloride	3.964	84	9238	0.248	ppbv #	85
32) 1,1,1-Trichloroethane	6.021	97	7831m	0.069	ppbv	
35) Carbon Tetrachloride	6.510	117	6833	0.056	ppbv	92
52) Tetrachloroethene	10.623	164	4430m	0.071	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102522\
Data File : VL039806.D
Acq On : 26 Oct 2022 9:01
Operator : SY/AP
Sample : QC102422 CAN 10279
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC102422 CAN 10279

Quant Time: Oct 28 08:41:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Oct 24 20:00:55 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/31/2022
Supervised By :Mahesh Dadoda 10/31/2022

