

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL102721\
 Data File : VL037916.D
 Acq On : 27 Oct 2021 14:56
 Operator : SY/AP
 Sample : M4377-02 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 VMP-5-IMPLANT

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/28/2021
 Supervised By : Mahesh Dadoda 10/29/2021

Quant Time: Oct 28 06:41:19 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1460122	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.19	114	4195785	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.10	117	3546220	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2405931	9.70	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.00%

Target Compounds					Ovalue
13) Ethanol	3.64	45	86282m	10.220	ppbv
15) Acetone	3.89	43	66564m	0.508	ppbv
19) Isopropyl Alcohol	4.01	45	59015m	0.724	ppbv
20) Methylene Chloride	4.37	84	50539m	0.763	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL102721\
Data File : VL037916.D
Acq On : 27 Oct 2021 14:56
Operator : SY/AP
Sample : M4377-02 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VMP-5-IMPLANT

Quant Time: Oct 28 06:41:19 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL100421
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/28/2021
Supervised By :Mahesh Dadoda 10/29/2021

