

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL102721\
 Data File : VL037922.D
 Acq On : 27 Oct 2021 19:01
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/28/2021
 Supervised By :Mahesh Dadoda 10/29/2021

Quant Time: Oct 28 07:07:32 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1226034	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	7.19	114	3740266	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3120483	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2163184	9.91	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	12509m	0.144	ppbv	
20) Methylene Chloride	4.36	84	26084m	0.469	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL102721\
Data File : VL037922.D
Acq On : 27 Oct 2021 19:01
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Quant Time: Oct 28 07:07:32 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL100421
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 10/28/2021
Supervised By :Mahesh Dadoda 10/29/2021

