

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
 Data File : VL026264.D
 Acq On : 29 Oct 2015 11:57
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Quant Time: Oct 29 13:56:22 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 29 12:22:38 2015
 QLast Update : Thu Oct 29 12:22:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.69	49	582577	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.37	114	1607167	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.79	117	1863017	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.33	95	1435312	10.07	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.92	62	4849	0.11	ppbv	100
32) 1,1,1-Trichloroethane	7.63	97	13140	0.11	ppbv	99
35) Carbon Tetrachloride	8.20	117	12098	0.10	ppbv	96
38) Trichloroethene	9.10	130	8912	0.11	ppbv	94
52) Tetrachloroethene	12.84	164	9463	0.10	ppbv	94
63) 1,1,2,2-Tetrachloroethane	15.53	83	21908	0.13	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
Data File : VL026264.D
Acq On : 29 Oct 2015 11:57
Operator : SY/AP
Sample : VSTDICC0.1
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Quant Time: Oct 29 13:56:22 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 29 12:22:38 2015
QLast Update : Thu Oct 29 12:22:38 2015
Response via : Initial Calibration

