

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
 Data File : VL026265.D
 Acq On : 29 Oct 2015 12:35
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Mmdadoda
 10/30/2015 12:06:37 PM

Quant Time: Oct 29 14:27:15 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 29 14:27:15 2015
 QLast Update : Thu Oct 29 13:56:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.68	49	572206	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.36	114	1583146	10.00	ppbv	-0.01
55) Chlorobenzene-d5	13.79	117	1825050	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.33	95	1417390	10.22	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.92	62	1896	0.02	ppbv	# 65
32) 1,1,1-Trichloroethane	7.63	97	5588	0.02	ppbv	98
35) Carbon Tetrachloride	8.20	117	4979	0.02	ppbv	98
38) Trichloroethene	9.10	130	3994	0.02	ppbv	# 82
52) Tetrachloroethene	12.83	164	4089	0.02	ppbv	# 87
63) 1,1,2,2-Tetrachloroethane	15.54	83	8725m	0.02	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
Data File : VL026265.D
Acq On : 29 Oct 2015 12:35
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Mmdadoda
10/30/2015 12:06:37 PM

Quant Time: Oct 29 14:27:15 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 29 13:56:09 2015
Response via : Initial Calibration

