

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
 Data File : VL026272.D
 Acq On : 29 Oct 2015 19:42
 Operator : SY/AP
 Sample : G4169-02DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA1DL

Quant Time: Oct 30 02:16:10 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Oct 29 14:29:54 2015
 QLast Update : Thu Oct 29 14:29:54 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.68	49	599807	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.36	114	1652190	10.00	ppbv	-0.01
55) Chlorobenzene-d5	13.78	117	1912646	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.32	95	1500360	10.34	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.59	51	122443	1.62	ppbv	99
13) Ethanol	4.39	45	1455	0.27	ppbv #	33
15) Acetone	4.66	43	41882	0.53	ppbv #	62
20) Methylene Chloride	5.18	84	7108	0.21	ppbv	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
Data File : VL026272.D
Acq On : 29 Oct 2015 19:42
Operator : SY/AP
Sample : G4169-02DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA1DL

Quant Time: Oct 30 02:16:10 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 29 14:29:54 2015
QLast Update : Thu Oct 29 14:29:54 2015
Response via : Initial Calibration

