

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
 Data File : VL026277.D
 Acq On : 29 Oct 2015 23:16
 Operator : SY/AP
 Sample : G4169-05DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Quant Time: Oct 30 04:25:50 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Oct 29 14:29:54 2015
 QLast Update : Thu Oct 29 14:29:54 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.67	49	589682	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	8.35	114	1622671	10.00	ppbv	-0.02
55) Chlorobenzene-d5	13.77	117	1909698	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.31	95	1488412	10.27	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.58	51	36156	0.49	ppbv	96
9) Propene	3.60	41	12181	0.34	ppbv #	67
13) Ethanol	4.36	45	1692	0.32	ppbv #	33
15) Acetone	4.63	43	167826	2.15	ppbv	94
20) Methylene Chloride	5.17	84	6039	0.18	ppbv	95
26) Hexane	6.68	57	9887	0.12	ppbv #	43
34) 2-Butanone	6.24	43	14695	0.15	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL102915\
Data File : VL026277.D
Acq On : 29 Oct 2015 23:16
Operator : SY/AP
Sample : G4169-05DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Quant Time: Oct 30 04:25:50 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL102915AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 29 14:29:54 2015
QLast Update : Thu Oct 29 14:29:54 2015
Response via : Initial Calibration

