

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL103018\
 Data File : VL032704.D
 Acq On : 30 Oct 2018 18:44
 Operator : SY/AP
 Sample : J5702-01DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 FIA-1DDL2

Manual Integrations
 APPROVED

MMDadoda
 10/31/2018 10:40:03 AM

Quant Time: Oct 31 05:51:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 25 05:15:43 2018
 QLast Update : Thu Oct 25 05:15:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1327788	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3205439	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3240958	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.59	95	2407057	9.97	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.02	51	309715	2.983	ppbv	100
9) Propene	3.06	41	84505m	1.978	ppbv	
13) Ethanol	3.67	45	11398m	0.933	ppbv	
15) Acetone	3.91	43	1151142	9.389	ppbv	98
19) Isopropyl Alcohol	4.06	45	98518	1.128	ppbv #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL103018\
Data File : VL032704.D
Acq On : 30 Oct 2018 18:44
Operator : SY/AP
Sample : J5702-01DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
FIA-1DDL2

Manual Integrations
APPROVED

MMDadoda
10/31/2018 10:40:03 AM

Quant Time: Oct 31 05:51:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 25 05:15:43 2018
Response via : Initial Calibration

