

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL103018\
 Data File : VL032708.D
 Acq On : 30 Oct 2018 21:17
 Operator : SY/AP
 Sample : J5693-01DL2 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-1DL2

Manual Integrations
 APPROVED

MMDadoda
 10/31/2018 10:40:08 AM

Quant Time: Oct 31 06:56:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 26 06:56:27 2018
 QLast Update : Thu Oct 25 05:15:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1315916	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3298813	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3262424	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.59	95	2402630	9.89	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.05	41	6129m	0.145	ppbv	
13) Ethanol	3.68	45	14322m	1.183	ppbv	
15) Acetone	3.93	43	116495	0.959	ppbv	97
26) Hexane	5.79	57	16095m	0.122	ppbv	
36) Benzene	7.02	78	27631m	0.103	ppbv	
52) Tetrachloroethene	11.39	164	15859m	0.170	ppbv	
53) Toluene	9.96	91	128707	0.419	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL103018\
Data File : VL032708.D
Acq On : 30 Oct 2018 21:17
Operator : SY/AP
Sample : J5693-01DL2 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-1DL2

Manual Integrations
APPROVED

MMDadoda
10/31/2018 10:40:08 AM

Quant Time: Oct 31 06:56:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 25 05:15:43 2018
Response via : Initial Calibration

