

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110121\
 Data File : VL037944.D
 Acq On : 1 Nov 2021 18:55
 Operator : SY/AP
 Sample : QC102721 CAN 10604
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC102721 CAN 10604

Quant Time: Nov 02 06:07:22 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 02 05:54:07 2021
 QLast Update : Tue Nov 02 05:54:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.658	49	1479201	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.163	114	4276959	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.073	117	3635385	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.400	95	2593238	9.813	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110121\
Data File : VL037944.D
Acq On : 1 Nov 2021 18:55
Operator : SY/AP
Sample : QC102721 CAN 10604
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC102721 CAN 10604

Quant Time: Nov 02 06:07:22 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 02 05:54:07 2021
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

