

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110221\
 Data File : VL037964.D
 Acq On : 2 Nov 2021 11:02
 Operator : SY/AP
 Sample : VL1102ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1102ABL01

Quant Time: Nov 03 03:49:30 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 02 05:54:07 2021
 QLast Update : Tue Nov 02 05:54:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.655	49	1339510	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.156	114	3935308	10.000	ppbv	-0.02
55) Chlorobenzene-d5	12.063	117	3366462	10.000	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.391	95	2362857	9.656	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110221\
Data File : VL037964.D
Acq On : 2 Nov 2021 11:02
Operator : SY/AP
Sample : VL1102ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1102ABL01

Quant Time: Nov 03 03:49:30 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 02 05:54:07 2021
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

