

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110221\
 Data File : VL037969.D
 Acq On : 2 Nov 2021 15:51
 Operator : SY/AP
 Sample : M4447-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 11/05/2021
 Supervised By : Mahesh Dadoda 11/09/2021

Quant Time: Nov 03 03:50:07 2021

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Nov 03 03:50:07 2021

QLast Update : Tue Nov 02 05:54:07 2021

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1365661	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.163	114	3814286	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.069	117	3439827	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.397	95	2418602	9.673	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.700%
Target Compounds						
						Qvalue
9) Propene	3.025	41	14903	0.163	ppbv	95
13) Ethanol	3.607	45	378818m	41.195	ppbv	
15) Acetone	3.867	43	59719m	0.529	ppbv	
19) Isopropyl Alcohol	3.967	45	511282m	7.085	ppbv	
20) Methylene Chloride	4.337	84	96347	1.555	ppbv	93
26) Hexane	5.677	57	30146	0.190	ppbv #	41
38) Trichloroethene	7.822	130	559714	4.341	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110221\
Data File : VL037969.D
Acq On : 2 Nov 2021 15:51
Operator : SY/AP
Sample : M4447-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/05/2021
Supervised By :Mahesh Dadoda 11/09/2021

Quant Time: Nov 03 03:50:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Nov
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

