

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110221\
 Data File : VL037971.D
 Acq On : 2 Nov 2021 17:09
 Operator : SY/AP
 Sample : M4447-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/05/2021
 Supervised By :Mahesh Dadoda 11/09/2021

Quant Time: Nov 03 03:50:21 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LTue Nov 03 03:50:21 2021
 QLast Update : Tue Nov 02 05:54:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.658	49	1338385	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.163	114	3743353	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.066	117	3237471	10.000	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.397	95	2309320	9.813	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.100%
Target Compounds						
					Qvalue	
13) Ethanol	3.626	45	33265m	3.691	ppbv	
15) Acetone	3.867	43	84857m	0.767	ppbv	
19) Isopropyl Alcohol	4.005	45	11400m	0.161	ppbv	
20) Methylene Chloride	4.340	84	136459	2.248	ppbv	96
26) Hexane	5.681	57	46865	0.301	ppbv #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110221\
Data File : VL037971.D
Acq On : 2 Nov 2021 17:09
Operator : SY/AP
Sample : M4447-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-1DL

Quant Time: Nov 03 03:50:21 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/05/2021
Supervised By : Mahesh Dadoda 11/09/2021

