

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110320\
 Data File : VL035817.D
 Acq On : 3 Nov 2020 23:56
 Operator : SY/AP
 Sample : L4604-01DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

MMDadoda
 11/5/2020 2:14:24 PM

Quant Time: Nov 04 05:55:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1208693	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	5242840	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	4773354	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	3405285	10.04	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.00	41	8040m	0.152	ppbv	
15) Acetone	3.85	43	22775	0.176	ppbv #	40
20) Methylene Chloride	4.32	84	22257m	0.245	ppbv	
29) Chloroform	5.73	83	31492	0.114	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110320\
Data File : VL035817.D
Acq On : 3 Nov 2020 23:56
Operator : SY/AP
Sample : L4604-01DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

MMDadoda
11/5/2020 2:14:24 PM

Quant Time: Nov 04 05:55:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

