

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110718\
 Data File : VL032717.D
 Acq On : 7 Nov 2018 14:58
 Operator : SY/AP
 Sample : J5809-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LIHA-IA-DUP-110118DL

Manual Integrations
 APPROVED

MMDadoda
 11/12/2018 12:30:36 PM

Quant Time: Nov 08 12:50:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 25 2018
 QLast Update : Thu Oct 25 05:15:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1531319	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3853692	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3730304	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2822553	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.06	41	13443m	0.273	ppbv	
13) Ethanol	3.66	45	1911641	135.713	ppbv	98
15) Acetone	3.92	43	216104m	1.528	ppbv	
17) tert-Butyl alcohol	4.37	59	21827m	0.525	ppbv	
19) Isopropyl Alcohol	4.04	45	147359	1.463	ppbv	100
53) Toluene	9.96	91	55969m	0.156	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110718\
Data File : VL032717.D
Acq On : 7 Nov 2018 14:58
Operator : SY/AP
Sample : J5809-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LIHA-IA-DUP-110118DL

Manual Integrations
APPROVED

MMDadoda
11/12/2018 12:30:36 PM

Quant Time: Nov 08 12:50:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 25 05:15:43 2018
Response via : Initial Calibration

