

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110718\
 Data File : VL032736.D
 Acq On : 8 Nov 2018 3:41
 Operator : SY/AP
 Sample : J5823-06DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-02DL

Manual Integrations
 APPROVED

MMDadoda
 11/12/2018 12:30:25 PM

Quant Time: Nov 12 10:15:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct 25 2018
 QLast Update : Thu Oct 25 05:15:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1333077	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.30	114	3312126	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.25	117	3235789	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2493071	10.35	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	19987	0.466	ppbv	93
13) Ethanol	3.64	45	898758	73.294	ppbv	96
15) Acetone	3.90	43	1503030	12.210	ppbv	97
19) Isopropyl Alcohol	4.03	45	72756	0.830	ppbv #	1
36) Benzene	7.01	78	79598	0.296	ppbv #	91
44) 2,2,4-Trimethylpentane	7.99	57	89732m	0.195	ppbv	
53) Toluene	9.95	91	1213864	3.937	ppbv	97
58) Ethyl Benzene	12.87	91	45775m	0.122	ppbv	
59) m/p-Xylene	13.14	91	163372m	0.488	ppbv	
60) o-Xylene	13.86	91	52494m	0.163	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110718\
Data File : VL032736.D
Acq On : 8 Nov 2018 3:41
Operator : SY/AP
Sample : J5823-06DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-02DL

Manual Integrations
APPROVED

MMDadoda
11/12/2018 12:30:25 PM

Quant Time: Nov 12 10:15:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL101818AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Oct
QLast Update : Thu Oct 25 05:15:43 2018
Response via : Initial Calibration

